

Performance data sheet

Testnaam	Kobalt [mol/volume] in serum of plasma
LOINC code	25378-1
Detection limit	<4 nmol/l
Imprecision	Low 14,1 nmol/l: CV% 4,0 High 103,0 nmol/l: CV% 2,7
Measurement range	<4 - 903 nmol/l
Traceability	Atomic Spectroscopy Standard Cobalt 1000 mg/l in 2% HNO ₃ , Perkin Elmer Interne standaard Itrium, Perkin Elmer