

Performance data sheet

Testnaam	Koper [mol/tijd] in 24-uurs urine
LOINC code	14666-2
Detection limit	0,12 umol/l
Imprecision	Low 0,35 umol/l: CV% 4,5 High 5,2 umol/l: CV 3,4
Measurement range	0,12 - 88,6 umol/l
Traceability	Atomic Spectroscopy Standard Koper 1000 mg/l, Perkin Elmer Interne standaard Rhodium, Perkin Elmer