

# Performance data sheet

Testnaam	Seleen [mol/volume] in serum of plasma
LOINC code	25521-6
Detection limit	<0,16 umol/l
Imprecision	Low 1,03 umol/l: CV% 4,7 High 2,12 umol/l: CV% 2,8
Measurement range	<0,16 - 6,37 umol/l
Traceability	Atomic Spectroscopy Standard Seleem 1000 mg/l, Perkin Elmer Interne standaard Germanium, Perkin Elmer