

Performance data sheet

Testnaam	Nikkel [mol/volume] in serum of plasma
LOINC code	25486-2
Detection limit	15,0 nmol/l
Imprecision	Low 14,4 nmol/l: CV% 14,8 High 111,4 nmol/l: CV% 5,7
Measurement range	15,0 - 912 nmol/l
Traceability	Atomic Spectroscopy Standard Nikkel 1000 mg/l, Perkin Elmer Interne standaard Ittrium, Perlin Elmer