

# Performance data sheet

Testnaam	IJzer [mol/tijd] in 24-uurs urine
LOINC code	14799-1
Detection limit	0,86 umol/l
Imprecision	Low 3,3 umol/l: CV% 7,7 High 27,9 umol/l: CV% 5,0
Measurement range	0,86 - 49,6 umol/l
Traceability	Atomic Spectroscopy Standard IJzer 1000 mg/l, Perkin Elmer Interne standaard Itrium, Perkin Elmer