

# Performance data sheet

**Testnaam** Proteïne S.vrij Ag gemeten/normaal in plaatjesarm plasma  
m.b.v. immunologische methode

---

LOINC code 27821-8

Detection limit 10 %

Imprecision  
Within run:  
Low 27%: CV% 5,3  
High 87%: CV% 2,4  
Between run:  
Low 29%: CV% 4,8  
High 88%: CV% 3,0

Measurement range 10 - 150 %

Traceability Calibrator van Siemens