

Performance data sheet

Testnaam	5-Hydroxyindolacetaat/kreatinine [molaire ratio] in 24-uurs urine
-----------------	--

LOINC code	47545-9
Detection limit	0,38 umol/l
Imprecision	Low 17,42 umol/l: CV% 4,6 High 72,57 umol/l: CV% 4,3
Measurement range	volgt
Traceability	SKML standaard VM01