

Performance data sheet

Testnaam	Homocysteïne [mol/volume] in serum of plasma
LOINC code	13965-9
Detection limit	<1,0 umol/l
Imprecision	Within run: Low 9,3 umol/l: CV% 1,94 Mid 52,6 umol/l: CV% 2,91 High 110,5 umol/l: CV% 2,06 Between run: Low 8,9 umol/l: CV% 5,95 Mid 51,7 umol/l: CV% 3,81 High 108,6 umol/l: CV% 5,31
Measurement range	<1,0 - >190,0 umol/l
Traceability	Grondstof Interne standaard Homocystine-d8 (C8H8D8N2O4S2 MW 276,4) DLM-3619-0/ 100 mg/ Cambridge Isotopic laboratorie Grondstof D,L-Homocystine (C8H16N2O4S2 Mw 268,4) H0501/ 10 g/ Sigma