

Performance data sheet

Testnaam	Homovanillaat/kreatinine [molaire ratio] in 24-uurs urine
----------	---

LOINC code	25931-7
Detection limit	0,43 umol/l
Imprecision	Low 6,19 umol/l: CV% 9,7 High 11,30 umol/l: CV% 7,7
Measurement range	volgt
Traceability	Cerilliant (Sigma - aldrich C-111) 1 mg /ml