

# Performance data sheet

Testnaam	Vanillylmandelaat/kreatinine [molaire ratio] in 24-uurs urine
----------	---

---

LOINC code	26004-2
Detection limit	0,30 umol/l
Imprecision	Low 7,59 umol/l: CV% 6,0 High 43,01 umol/l: CV% 3,2
Measurement range	volgt
Traceability	Cerilliant (Sigma - aldrich C-111) 1 mg /ml