

Performance data sheet

Testnaam Insuline [eenheden/volume] in serum of plasma

LOINC code	20448-7
Detection limit	0,6 mU/l
Imprecision	Low 6,2 mU/l CV% 4,7 Medium 98 mU/l CV% 4,2 High 172 mU/l CV% 4,0
Measurement range	0,6 - 400 mU/l
Traceability	Standaard, WHO NIBSC 1st IRP 66/304