

Performance data sheet

Testnaam Pancreas elastase [massa/massa] in feces

LOINC code	25907-7
Detection limit	50 ug/g faeces
Imprecision	Low 100 ug/g: CV% 7,7 High 200 ug/g: CV% 7,9
Measurement range	50 - 500 ug/g faeces
Traceability	Standaard, Bio Serve Diagnostics